

CERTIFICATE OF COMPLIANCE

Certificate Number 2018-12-19; 2019-08-23 (A0/C1); 2020-04-01(A0/C2); 2021-06-25 (A1/C2)-E340287
Report Reference E340287-D1000-1/A1/C2-UL
Date 2018-12-19; 2019-08-23 (A0/C1); 2020-04-01(A0/C2); 2021-06-25 (A1/C2)
Issued to: Knick Elektronische Messgeraete GmbH & Co. KG
Applicant Company: Beuckestrasse 22
Berlin, 14163 Germany
Listed Company: Same as Applicant

**This is to certify that
representative samples of**

High-Voltage isolation amplifiers
VariTrans P2900x P2/00 or VariTrans P2900x P2/01, whereby
"X" can be "0" or "1" not safety relevant.
BL590; BL591


Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety: UL 61010-1, 3rd Edition, May 11, 2012, Revised April 29 2016,
CAN/CSA-C22.2 No. 61010-1-12, 3rd Edition, Revision dated
April 2016
Additional Standards: IEC 61010-2-030: 2010 (First Edition), UL/CSA 61010-2-030.
Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information.

This Certificate of Compliance does not provide authorization to apply the UL Mark. Only the UL Follow-Up Services Procedure provides authorization to apply the UL Mark.

Only those products bearing the UL Mark should be considered as being UL Certified and covered under UL's Follow-Up Services.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

